

Chargenzertifikat

Bezeichnung **KCl 3mol/l**
Artikel Nr.: **10.30.123x**
Charge: **C1A11**

Ungeöffnet haltbar bis: **17.05.2025**

Analysenwerte:

Prüfmerkmale	Prüfmethode	Soll - Wert	Ist - Wert
Aussehen	visuell	wasserklar	entspricht
Gehalt	Grafimetrisch	223,65 g/l KCl,	223,65 g/l KCl,

Es wird bestätigt, dass die Lieferung den Vereinbarungen bei der Bestellannahme entspricht.

Unterschrift:



17.05.2022

Angelika Christein

Dieses Zertifikat wurde sorgfältig nach bestem Wissen erstellt. Es entbindet jedoch nicht zur Wareingangskontrolle und begründet keine Ansprüche Dritter an die es weitergeleitet wird. Eine Eigenschaftszusicherung im rechtlichen Sinne ist hiermit nicht verbunden.

Certificate of analysis

Product: KCl 3mol/l
cat. no.: 10.30.123x
Batch no.: C1A11
Minimum shelf life: 17.05.2025

Analysis:

Characteristic	Method	Target	Actual
Appearance	visual	clear	corresponds
Concentration	weighted	223,65 g/l KCl,	223,65 g/l KCl,

Person in Charge:



17.05.2022

Angelika Christein

This certificate was carefully prepared after the best knowledge. It, however, doesn't release for receipt of goods control.
A feature assurance in the legal meaning isn't connected hereby.